

**OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC HEALTH
OFFICER, PURI**

BID DOCUMENT

FOR

EQUIPMENT / INSTRUMENTS & PRINTING, PURI

Reference No.

Date: 05.09.2020

Address for communication with Email ID:

O/O- CDM & PHO, PURI, PIN-752001

Website: www.puri.nic.in, E-Mail: dpmupuri@gmail.com

CONTENT

Section	Clause	Component	Page No.
Section – I General Information About Tender	1	Background	5
	2	Description & Directives	5
Section – II Specific Information for Bidders	1	Eligibility Criteria	8
	2	Tenure of the Contract	8
	3	Important Notes for Bidders	9
	4	Bidding Process	9
	5	Evaluation and Selection	11
	6	Award of Contract	11
	7	Performance Security	11
Section – III Conditions of Contract	1	Contract Management	13
	2	Payment Terms	15
	3	Resolution of Disputes	15
Annexure	I	List of Equipments & Aids (Equipment / Instruments & Printing, Puri)	16
	II	Format for Turnover Certificate	17
	III	Format for Declaration	18
	IV	Format for General Details of the Bidder	19
	V	Agreement Format	20
	VI	Format of Bank Guarantee for Performance Security	22

NOTICE INVITING BID

Office of the Chief District Medical & Public Health Officer, Puri

Website: www.puri.nic.in, E-mail: dpmupuri@gmail.com

Bid Reference No.:

Date: 05/09/2020

Bids are invited from eligible firms/company/agency/suppliers/distributers to supply the **Equipment / Instruments & Printing, Puri** as per the details in **Annexure-I** and as per the particulars mentioned below:

Sl. No.	Particulars	Date and Time	
1.	Date of release of bid	05.09.2020	
2	Last date & time for submission of bid	23.09.2020, 5 pm	Office of the CDM & PHO, PURI
3.	Date & time of Technical bid opening	24.09.2020, 11 am	Venue: Conference Hall, Office of the CDM & PHO, PURI
4.	Date of opening of Price Bid		

The bid document with all information relating to the bidding process including the cost of bid document, EMD, Eligibility criteria and terms & conditions are available in the district website: www.puri.nic.in. The last date of submission of bid is on or before **23.09.2020 by 5 pm** to the O/o- CDM&PHO, PURI through Speed Post/Regd. Post/Courier only. The bid document should be superscribed with "**Equipment / Instruments & Printing, Puri**". The Authority reserves the right to accept/reject any part thereof or all the bids without assigning any reason thereof.

-Sd-

CDM & PHO, Puri

SECTION – I

GENERAL INFORMATION ABOUT TENDER

1. Background

Chief District Medical & Public Health Officer, Puri is inviting bids to supply **Equipment / Instruments & Printing, Puri** (see **Annexure-I** for details). This agreement describes the terms and conditions under which the agency (successful bidder to whom contract is awarded) agrees upon to supply the Equipments.

Note:

All products shall be owned by the Chief District Medical & Public Health Officer, PURI and provided to the agency (successful bidder to whom contract is awarded) to supply the Equipments under the terms and conditions of this contract.

2. DESCRIPTION & DIRECTIVES

Chief District Medical & Public Health Officer, PURI [hereinafter referred as Tender Inviting Authority unless the context otherwise requires] invites sealed bids for supplying Equipments to ~~Delivery Points~~ as per **Annexure-I(Equipment / Instruments & Printing , Puri)**

3. OTHER RELEVANT INFORMATION

- a) Tender Type : Manual Tender, 2 envelopes System - one labeled as Technical Bid and the other one as Financial Bid.
- b) Mode of Operation : To be decided by the Chief District Medical & Public Health Officer, Puri .
- c) Nature of work : Supply of Equipments
- d) Cost of Tender Paper : **Rs. 2000/-** (Rupees two thousand, inclusive of GST) in the form of Demand Draft/ Pay Order in favour of **ZSS NON NRHM, PURI** , payable at PURI from any Nationalized/ Scheduled Bank. (Tender paper can be downloaded from district website www.puri.nic.in and should be enclosed as cost of tender paper).

f) EMD (Refundable) : **Rs.10,000/-** (Rupees ten thousand) in form of Demand Draft/ Pay Order in favour of **ZSS NON NRHM, PURI** , payable at PURI from any Nationalized / Scheduled Bank.

g) Important Dates : Tender Release:05.09.2020

Last date & time for bid submission: 23.09.2020 by 05 PM to the

O/o-CDM & PHO, PURI through Speed Post/Regd.
Post/ Courier only.

Bid Opening: 24.09.2020, 11.00 AM in Conference Hall, Office of the

CDM & PHO, PURI

h) Validity of Bid : 180 days from the date of bid submission

i) Address for Communication: **Office of the CDM & PHO, PURI , PIN -752001**

E-mail: dpmupuri@gmail.com

Note:

The bidders shall be solely responsible for checking the website www.puri.nic.in for any addendum/amendment issued subsequent to the release of bid document and take into consideration the same while preparing and submitting the bids.

SECTION – II

SPECIFIC INFORMATION FOR BIDDERS

1. ELIGIBILITY CRITERIA

The interested Bidders shall have to comply to the following criteria to participate in the tendering process –

- a. Authorization certificate in case of Authorized distributor/ Manufacturing license in case of Manufacturer/ Import license in case of Importer.
- b. Must have all relevant statutory documents such as Registration Certificate of firm, PAN Card, GST Registration Certificate along with the following:
 - i. Valid ISO certificate if any
 - ii. Valid Drug License in case of supply of Drugs.
 - iii. CE /US FDA/ IEC certificates of the manufacturer as per technical specification if any
- c. Financial Statement of last three financial year duly certified by Chartered Accountant. and the average annual turnover should not be less than Rs. 03.00 Crore of last 3 conjugative year with audit report. (Report on the financial standing of the Bidder such as Turnover Statement, duly certified by a Chartered Accountant, as in Annexure-II for the Financial Years 2016-17, 2017-18 & 2018-19 to be furnished. Provisional Turnover shall not be considered.)
- d. End users certificate should be issued from concern authority of the different institutions for last five years in name of distributor. Again the applicant should be submitted the END users/Market standing certificate from the concern authority that the distributor has supplying them the medical instruments, equipments ,BMW articles and bedding clothing etc.
- e. The bidder should have minimum 5 years experience of the required items which is mention in end users certificate from the concern authority. Minimum 10 nos end users certificate from last five years with 5 to 10 nos order copies of each year for last 5 years. The bidder should be a GST registered in state of odisha only.
- f. Income Tax Clearance certificate for the last three FY.
- g. Performance certificate of last three years supply to Govt. institution along with Order copy. Without order copy performance certificate is not valid.
- h. Sample or Technical Brochure of the items quoted. There are few items having no technical specification (prototype enclosed) which will be finalized by the Purchase committee after verification of samples or technical brochures.
- i. Should neither be convicted by any court of law nor be blacklisted/ debarred on financial/ administrative/ techno-legal ground by any appropriate authority. A declaration regarding this should be submitted in a Rs. 10/- stamp paper in the format prescribed at **Annexure III**. The bidder should not be involved in any criminal or civil suit. If the bidder is involved in any criminal or civil suit, then details of all such suits should be submitted. The Tender Inviting Authority reserves the right to consider/ not consider any offer based on the gravity/ implication of suits pending against the bidder.
- j. The Original Tender Booklet with Conditions and the schedules signed by the bidder at the bottom of each page with his official seal duly affixed.

2. TENURE OF THE CONTRACT

- a. The contract shall be awarded for a period of **one year** from the date of signing of contract.
- b. In the event of the contract being terminated by the Tender Inviting Authority before the completion of the contract period for reasons of unsatisfactory performance or breach of contractual conditions, the Performance Security of the agency shall be liable to be forfeited.

3. IMPORTANT NOTES FOR BIDDERS:

- a. EMD will be refunded to the unsuccessful bidders once the successful bidder submits the Performance Security and signs the contract.
- b. EMD will be refunded to the successful bidder after submission of the required Performance Security.
- c. In case of unwillingness of L1 bidder to enter into contract or bid withdrawal in any form, the EMD submitted by the bidder will be liable for forfeiture and the Tender Inviting Authority will have the right to proceed for a contract with the next lowest responsive bidder.

4. BIDDING PROCESS

The bidder should submit its proposal in two envelopes - "Envelope A" and "Envelope B". Envelope A should include qualification documents and Envelope B should contain financial offer. Documents to be submitted in the respective envelopes are as follows:-

A. TECHNICAL BID – ENVELOPE A: (Both)

The Bidders have to submit following documents in their technical bids with page references-

Sl. No.	Particulars	Page No.
1.	Cost of Tender Paper	
2.	EMD	
3.	Self-attested copy of Authorization certificate in case of Authorized distributor/ Manufacturing license in case of Manufacturer/ Import license in case of Importer	
4.	Self-attested copy of GST Registration Certificate	
5.	Self-attested copy of PAN card	
6.	Turnover Certificate as per Annexure-II	
7.	Income Tax Clearance certificate for the last three FY	

8.	Performance certificate of last five years supply to Govt. Institution along with Order copy	
9.	User certificate from the concerned Institution for the last three year	
10.	Declaration as per Annexure-IV	
11.	General Details of the Bidder as per Annexure-V	
12.	Bank Details of the bidder as follows – a. Name of the Bank and Branch b. Type of Account - Savings/ Current c. Bank Account No. d. IFS Code	

LIST OF ITEM(S) QUOTED (TO BE ENCLOSED IN ENVELOPE FOR TECHNICAL BID)

Sl. No.	Name of Item(s)	Category (Equipment / Instruments & Printing , Puri)	Name of Manufacturer	Make	Model Name	*Details of offered product at PageNo.(s)
1	Head Light	<p>Purpose: Coaxial luminaire provides shadow free illumination for improved efficiencies</p> <p>Light weight and comfortable head band options : Yes</p> <p>Supply of accessory like charger,cable & batteries : Yes</p> <p>Color of light : Cool day light</p> <p>Brightness (Lumen): 40</p> <p>Color Temperature (Degree Kelvin): 4500 degree K,</p> <p>Minimum LED Life (Hours): 20000</p> <p>Nature of focusing: Spot</p> <p>Portable compact design with no wires- into shirt pocket: Yes</p> <p>Charging time (Hours): 2-4Hours</p> <p>Color rendering index (CRI): 90</p> <p>Type of battery : Alkaline</p> <p>Run time (Minutes): 60</p> <p>ON/OFF control: Yes</p> <p>MINIMUM Spot diameter (mm): 10</p> <p>MAXIMUM Spot diameter (mm) : 60</p> <p>Standards:</p> <p>Product complies with: IEC 60601-1 and IEC60601-2-41 :Yes</p> <p>Model Should be CE/BIS certified :Yes</p> <p>Manufacturer should be ISO13485 certified: Yes</p>				
2	Ear Specula	<p>Shape: Oval cone design</p> <p>Material: Stainless Steel, antimagnetic, acid-resistant</p> <p>Supply:Set of 4nos.Diameter: 3mm, 4mm,5mm& 7mm respectively.</p> <p>Sterilization / Disinfection:Completely reusable and fully autoclavable.</p> <p>Each instrument shall be embossed/etched with manufacturer's name, initials or recognized trade-mark and the words 'Stainless Steel' or letters 'ss'</p> <p>Manufacturer should be ISO 13485 certified</p> <p>Product should be CE/BIS approved</p>				
3	Nasal Speculum	<p>Type: Thudichum</p> <p>Product: Nasal Specula</p> <p>Profile: Set of 3(Adult sizes-4&5 & pediatric sizes-3)</p> <p>Specialty: Rhinology Instruments</p> <p>Material: Stainless Steel(AS1410/420)</p> <p>Surface Finish: Satin</p> <p>Sterilization / Disinfection: Completely reusable and fully autoclavable.</p> <p>Each instrument shall be embossed/etched with manufacturer's name, initials or recognized trade-mark and the words 'Stainless Steel' or letters</p>				

		<p>'ss'</p> <p>Manufacturer should be ISO 13485 certified</p> <p>Product should be CE/BIS approved</p>				
4	Mouth Mirror	<p>Type Mirror With Handle</p> <p>Shape of Mirror: Round</p> <p>Diameter of Mirror (mm) set :18 millimeter & 20millimeter</p> <p>Material of housing of the mirror: Medical Grade Stainless Steel(316/ASi410)</p> <p>Anti Fog Coating Yes</p> <p>Thick Non-Slip Handle for perfect grip: Yes</p> <p>Mirror should be With minimum 10 Cm Length of Handle: Yes</p> <p>Mirror should be detachable from the handle: Yes</p> <p>Extra light and handy for more fatigue-free working : Yes</p> <p>GENERAL FEATURES</p> <p>Product Description: Dental Mouth Mirror or Dentist's Mirror</p> <p>Clinical Purpose: For allowing indirect vision by the dentist, reflecting light onto desired surfaces and retraction of soft tissues</p> <p>Disposable or Reusable: Reusable</p> <p>Sterile :Yes</p> <p>Autoclavable: Yes</p> <p>PACKAGING:Individually Packed</p> <p>CERTIFICATIONS & REPORTS</p> <p>Submission of Test Report from Central Govt/NABL/ILAC accredited Lab to prove the conformity to declared specification at the time of supply: YES</p> <p>Product certification: Manufacturer should be ISO13485 certified</p> <p>Product should be CE /IS approved</p> <p>WARRANTY & MAINTENANCE</p> <p>Warranty (Yrs) :Replacement warranty of 1year</p>				
5	Tongue Depressor	<p>MAYO TONGUE DEPRESSOR</p> <p>Types shape Flat with round ends</p> <p>Stainless Steel(ASI410/420)</p> <p>Surface Finish: Polish</p> <p>Sterilization / Disinfection: Completely reusable and fully autoclavable.</p> <p>Each instrument shall be embossed/etched with manufacturer's name, initials or recognized trade-mark and the words 'Stainless Steel' or letters 'ss'</p> <p>Manufacturer should be ISO 13485 certified</p> <p>Product should be CE/BIS approved</p> <p>Child Size 140mm</p> <p>Adult Size 200mm</p>				
6	Foreign Body	<p>Name: Alligator</p>				

	Removal Forceps	<p>Type: Forceps</p> <p>Size: 5 1/2"</p> <p>Surface Finish: Polish</p> <p>Curvature: Straight Jaws with angled Shanks</p> <p>Working Surface style: Horizontally Serrated</p> <p>Material: Stainless Steel(ASI410/420)</p> <p>Sterilization / Disinfection: Completely reusable and fully autoclavable.</p> <p>Each instrument shall be embossed/etched with manufacturer's name, initials or recognized trade-mark and the words 'Stainless Steel' or letters 'SS'</p> <p>Manufacturer should be ISO 13485 certified</p> <p>Product should be CE/BIS approved</p>				
7	Otoscope	<p>Diagnostic Otoscope rechargeable set is to be supplied in Leather / hard case preferably HIP (High Impact Plastic) with internal soft cushion material for easy portability and protection :Yes</p> <p>Light source : LED bulb</p> <p>Lamp Hour: 20,000 hours minimum /2years warranty of free replacement</p> <p>Number of Spare bulbs to be provided along with the otoscope:0</p> <p>Speculum adapter to be provided : Yes</p> <p>Number of sets of different sizes of Ear Specula to be provided along with otoscope : 2nos (One for Adult & one for pediatric)</p> <p>Disposable :No, It should be reusable</p> <p>Removable Otoscope head : Yes</p> <p>Rotatable lenses : Yes</p> <p>Rechargeable handle: Yes</p> <p>Provision for variable light intensity : Yes</p> <p>CRI: Minimum 90</p> <p>Color temperature: 4300K minimum</p> <p>CONSTRUCTION/FUNCTIONAL FEATURES</p> <p>Material of the device body: Medical Grade Stainless Steel/Medical grade ABS body</p> <p>Instrument Head Matt-black from inside: NA</p> <p>Fibre optics for Cool Light Transmission: Yes</p> <p>Magnification: 3x/3.5x</p> <p>Wide angle viewing lens allowing instrument under Magnification: Yes</p> <p>Provision of sealed system for Pneumatic otoscopy with detachable pneumatic bulb : NA</p> <p>Voltage Rating of the bulb (Volts)(Min) : 2.5/3/3.5</p> <p>Type of rechargeable battery: Li-ion/Ni-Cd/AA</p> <p>Capacity of the Battery (mAh):NA</p> <p>Battery Voltage(Volts) : 2.5/3/3.5</p> <p>Continuous runtime of the battery (Minutes):Minimum 60minutes</p> <p>Charger to be provided along with the device: Yes</p>				

		<p>Weight (gms):200-300</p> <p>STANDARD</p> <p>Conformity to Standard : EU-CE/USFDA/BIS</p> <p>Manufacturer Should be ISO13485 certified</p>				
8	Jobson Horn Probe	<p>Jobson Horne Probe size: 5.5"</p> <p>Features: double end one side screw and other side round loop.</p> <p>Rust free sturdy construction for better grip.</p> <p>A light-weight, flexible and strong.</p> <p>Probes designed as a dual-purpose instrument with Curette loop at one end and a threaded section at the other for holding wrapped cotton wool.</p> <p>Reduced patient trauma due to smooth edges.</p> <p>Sterilization / Disinfection: Completely reusable and fully autoclavable.</p> <p>Each instrument shall be embossed/etched with manufacturer's name, initials or recognized trade-mark and the words 'Stainless Steel' or letters 'ss'</p> <p>Manufacturer should be ISO 13485 certified</p> <p>Product should be CE/BIS approved</p>				
9	Tuning Fork	<p>Nickel-plated steel, material thickness 8 mm, Frequency: 128 Hz / 512 Hz and 1024 Hz.</p> <p>Sterilization / Disinfection: Completely reusable and fully autoclavable.</p> <p>Each instrument shall be embossed/etched with manufacturer's name, initials or recognized trade-mark and the words 'Stainless Steel' or letters 'ss'</p> <p>Manufacturer should be ISO 13485 certified</p> <p>Product should be CE/BIS approved</p>				
10	Noise Maker	<p>Bell made of brass. Minimum length -8.5cm and minimum weight - 70gram. Should be child friendly.</p> <p>Manufacturer should be ISO certified</p>				
11	Mouth mirror	<p>Type Mirror With Handle</p> <p>Shape of Mirror: Round</p> <p>Diameter of Mirror (mm) set :18 millimeter & 20millimeter</p> <p>Material of housing of the mirror: Medical Grade Stainless Steel(316/ASi410)</p> <p>Anti Fog Coating Yes</p> <p>Thick Non-Slip Handle for perfect grip: Yes</p> <p>Mirror should be With minimum 10 Cm Length of Handle: Yes</p> <p>Mirror should be detachable from the handle: Yes</p> <p>Extra light and handy for more fatigue-free working : Yes</p> <p>GENERAL FEATURES</p> <p>Product Description: Dental Mouth Mirror or Dentist's Mirror</p> <p>Clinical Purpose: For allowing indirect vision by the dentist, reflecting light onto desired surfaces and retraction of soft tissues</p>				

		<p>Disposable or Reusable: Reusable</p> <p>Sterile :Yes</p> <p>Autoclavable: Yes</p> <p>PACKAGING: Individually Packed</p> <p>CERTIFICATIONS & REPORTS</p> <p>Submission of Test Report from Central Govt/NABL/ILAC accredited Lab to prove the conformity to declared specification at the time of supply: YES</p> <p>Product certification: Manufacturer should be ISO13485 certified</p> <p>Product should be CE /IS approved</p> <p>WARRANTY & MAINTENANCE</p> <p>Warranty (Years) :Replacement warranty of 1year</p>				
12	Periodontal probe	<p>Double ended periodontal probe. Provides accurate pocket depth measurement – and information on sub gingival calculus and other root surface unevenness</p> <p>Ball end of ø 0,5 mm - Increases tactile sensitivity - Does not damage tissue or pierce pocket bottom</p> <p>Sterilization / Disinfection: Completely reusable and fully autoclavable.</p> <p>Each instrument shall be embossed/etched with manufacturer's name, initials or recognized trade-mark and the words 'Stainless Steel' or letters 'ss'</p> <p>Manufacturer should be ISO 13485 certified</p> <p>Product should be CE/BIS approved</p>				
13	Tooth Model	<p>Enlarged mouth model with toothbrush to demonstrate correct brushing and dental care.</p> <p>This model is three times life size and has normal anatomy. It's great for teaching proper brushing and flossing techniques.</p> <p>This teeth care demonstration model features a mouth with 32 teeth at three times life-size, for closer observation</p> <p>European CE Declaration according to 2014/53/EU and European Directive on Medical Devices</p> <p>Manufacturer's Quality Certification according to ISO 13485 Yes</p> <p>Packing Each model should be packed individually.</p> <p>Warranty 1 year</p>				
14	Torch	<p>Should provide even field of light</p> <p>Bright and natural white light by high performance LED.</p> <p>LED burns at least 20,000 hours.</p> <p>Metal /ABS casing.</p> <p>Easy to use with on/off switch.</p> <p>Includes blade holder, batteries and nylon case.</p> <p>Designed for durability, comes with a 2-year warranty including LED</p> <p>Adjustable spot size from 2-5CM</p> <p>Manufacturer should be ISO certified</p> <p>Product should be CE/BIS approved</p>				
15	Tooth Extraction Forceps Set(Adult)	GENERAL FEATURES				

		<p>Product description: Dental Forceps for extraction of tooth</p> <p>Purpose For dental extraction of tooth</p> <p>Utility Dental surgeries of Adult</p> <p>PRODUCT INFORMATION</p> <p>Items in Extraction forceps set: Upper anterior, Lower anterior, Lower molar (Right), Lower molar (Left), Upper molar (Left), Upper molar (right), upper premolar, lower premolar.</p> <p>anatomically-designed beaks for a sure, effective grip: Yes</p> <p>Thick and non-slip handle : Yes</p> <p>Longevity and corrosion resistant material: Yes stainless steel (ASI 410/420)</p> <p>Cleaning, disinfection and sterility: All instruments autoclavable</p> <p>DIMENSIONS & MATERIAL</p> <p>Minimum Handle length (Cm) :10</p> <p>Material of all instruments :Surgical Grade Stainless steel(ASI410/420)</p> <p>Rockwell hardness (HRC): 1</p> <p>Tensile strength (Mpa) : 1</p> <p>STANDARD</p> <p>Product conformity to standard ISO 13485</p> <p>Product should be CE/BIS approved</p> <p>PACKING MODE</p> <p>All the instruments along with user manual, warranty card, Dehumidifier, Cleaning brush, Cleaning cloth etc packed in a carrying case : Yes</p> <p>Packing Each carrying case packed Individually</p> <p>CERTIFICATIONS & REPORTS</p> <p>Submission of Test Report on form 39 or from Central Government /NABL/ILAC accredited Lab or Manufacture's in-house Test Report to prove the conformity to declared specification at the time of supply Yes</p> <p>Product conformity certificate is to be provided to the buyer at the time of supply Yes</p> <p>WARRANTY</p> <p>Minimum Warranty (yrs) : 5 years replacement</p>				
16	Dental Chair	Rate contract available with OSMCL. The purchase order to be placed by OSMCL once the site readiness for installation is received from the concern institutions.				
17	Dental X-ray with RVG	Rate contract available with OSMCL. The purchase order to be placed by OSMCL once the site readiness for installation is received from the concern institutions.				
18	Instrument Set	<p>As per requirement the clinician to decide the list of instruments</p> <p>Broad specification: Quality standard (Applicable to all surgical instruments):Instruments should be made up of stainless steel medical grade AISI 410 & 420. Test reports should be submitted in the technical bid.</p> <p>Part No. and the CE Mark must be engraved/Embossed on the instrument.</p>				

		<p>Sample demonstration</p> <p>Instrument should cover 2 years replacement warrantee.</p> <p>All the instruments should be autoclavable.</p>				
19	Haemoglobinometer (Strip/Cuvette)	Supply starts from OSMCL(provision under NVBDCP,IMR& MMR & DEIC)				
20	Glucometer	To be Supplied from OSMCL as per requirement of NCD.				
21	Colorimeter	<p>It is recommended to have Semi-Automated biochemistry analyzer in place of Colorimeter where the trained manpower (laboratory technicians are available).</p> <p>Rate contract available at OSMCL which is valid till 13/06/2020.</p>				
22	Binocular Microscope	<p>Microscope Type: laboratory binocular non-hinged type with built in light and with light intensity regulator</p> <p>Conformity to Indian Standard: Research IS:5204 latest</p> <p>ISI MARKED :NA</p> <p>Eye piece Type : Compensating</p> <p>Binocular Eye pieces Confirming to the requirements of IS: 8275/1976 (latest) :Yes</p> <p>Eye Piece with magnification : Set of Two For Binocular 10x</p> <p>Objective Type : achromatic</p> <p>Objective Magnification :100X,40X,10X,4X</p> <p>Numerical Apertures of Objective: 1.25</p> <p>Type of lamp for Illumination :LED</p> <p>Plano Concave mirror attachment :Yes</p> <p>Stage : Rectangular</p> <p>Size of Stage : 76 X 40 MM/Standards Size</p> <p>Coarse and Fine Movement of stage:Yes</p> <p>Co-axial focusing :Yes</p> <p>Micrometer Arrangement :No</p> <p>ACCESSORIES</p> <p>Extra eyepieces (inclusive in the scope of supply) : NA</p> <p>Carrying case (inclusive in the scope of supply):Yes</p> <p>WARRANTY</p> <p>Warranty 2 year</p> <p>CERTIFICATION</p> <p>Availability of Test Report from Central Govt/NABL/ILAC accredited lab to prove conformity to specification No</p> <p>Test Report No and Date N.A.</p> <p>Name and Address of the Lab N.A.</p> <p>Manufacturer Should be ISO13485 certified</p> <p>Model should be CE/USFDA /SI approved</p>				
23	Near Vision Chart	Name of Tool : Lea Symbols Near Vision Card/ISIHARA Near Vision Chart				

		<p>Type : "16 inches/40 centimeter Lea Near vision chart"</p> <p>Objective : To fulfill the criteria to be a good vision for children by measuring near visual acuity</p> <p>Use : "To assess the child's functional vision at near distances"</p> <p>Utility: Pre-school children of 3 to 5 years age group</p> <p>PRODUCT INFORMATION AND TESTING</p> <p>"Front side of the Lea symbol cards printed with proportionally spaced (log MAR) lines" : Yes</p> <p>"Back side of Lea symbol card printed with 25% and 50% spacing" : Yes</p> <p>"The card contains very familiar 4 symbols like circle, square, house and apple" : Yes</p> <p>All optotypes should be similar legibility : Yes</p> <p>Each line should have optotypes size ranging from 20/400 to 20/10 (6/120 to 6/3) equivalent : Yes</p> <p>Detection of mild amblyopia is possible with closed spacing of the symbols on Lea symbol card : Yes</p> <p>Test distance (cm) : 40</p> <p>Response key printed on the test card : Yes</p> <p>"Includes Student response or training card and instructions" : Yes</p> <p>Inclusive of conditioning flash card : Yes</p> <p>Can be used as wall hanging : Yes</p> <p>Material of the card "Non-tear waterproof material"</p> <p>Card size (cmxcm) : 20.3x25.4</p> <p>PACKING</p> <p>Packing (including user instructions) :Packed in a wallet</p> <p>CERTIFICATIONS</p> <p>Certification Available :Yes</p> <p>Manufacturer Quality certification (ISO)</p> <p>Product approval certificate for Clinical Use</p>				
24	Color Vision Chart	<p>Individuals with color vision defect should see a different figure from individuals with normal color perception.</p> <p>Makes use of the peculiarity that in red-green blindness, blue and yellow appear remarkably bright compared with red and green</p> <p>Diagnostic plates: intended to determine the type of color vision defect.</p> <p>International standard test chart books for color deficiency 38 plates with user manual</p> <p>Book should be specially printed to measure incrementally higher color sensitivity and specificity</p> <p>Manufacturer should be ISO certified</p>				
25	Snellen's Chart	<p>Snellen chart with Red/Green bar to test visual acuity.</p> <p>Assessment distance 6m.</p> <p>Size: 23 x 35.5 cm</p>				
26	Stadiometer	Standards:				

	<p>Conformity to Stadiometer as per ICDS standards : Yes</p> <p>Stadiometer designed to Measure Height of Adult and Children aged 24 Month and Above in Vertical Position : Yes</p> <p>Performance Parameters</p> <p>Smallest Graduation for Stadiometer (cm) : 0.1 cm</p> <p>Accuracy for Stadiometer (cm) : (+/-)0.1 cm</p> <p>Precision for Stadiometer (cm) : (+/-)0.1 cm</p> <p>Warranty (in Years) : 2</p> <p>Dimensional And Material Parameters</p> <p>Unit of Measurement : In Centimeters</p> <p>Measuring Length in vertical position of Stadiometer (should be able to measure up to 200 cm at least) : 200 centimeter</p> <p>Width of board for stadiometer (cm) : Approx. 5.5 cms</p> <p>Weight of the device (Kg) : 4 kilogram or less</p> <p>Material of Board : mild steel rust proof</p> <p>Additional Features</p> <p>No part Should be Loose or Shaking During Transportation : Yes</p> <p>Double sided graduation Parallel to Board for Easy Measurement : Yes</p> <p>Stable connection assuring precise and accurate measurement : Yes</p> <p>No need of calibration as all parts have prefixed position: Yes</p> <p>Stadiometer should have Firm flat surface : Yes</p> <p>Designed for heavy duty use in demanding conditions : Yes</p> <p>Simple and fast set up, no tool (screw driver etc) required : Yes</p> <p>Orientation of numerals on the measurement scale: parallel to the board : Yes</p> <p>Stadiometer should have large foot plate providing extra stable base and smoothly gliding measuring slide/wedge : Yes</p> <p>Stadiometer shall have Adjustable feet for stability on uneven or soft ground and also spirit level indicator to check flat surface : Yes</p> <p>Fold-up Mechanism and low-weight, making it compact and easy to transport : Yes</p> <p>Stadiometer shall be free from sharp edges and corners : Yes</p> <p>Stadiometer shall have friction feature between board and measuring slide/wedge for preventing the latter from drooping when released: Yes</p> <p>Durable, Resistance to excessive Humidity, High temperature resistant, water resistant and shock Resistant : Yes</p> <p>Stadiometer supplied with :</p> <p>i Carry bag or carry case made of long lasting fabric, zipper closer</p> <p>(ii) removable, adjustable shoulder/back strap(s) and</p> <p>(iii) small side pocket for storage of operating instructions, loose parts and/or stationery material : Yes</p> <p>Stadiometer supplied with Instructions for use, training, maintenance and trouble shooting in English, Hindi and any other regional language as per requirement, appropriately illustrated with pictograms : Yes</p> <p>Packed in a carton box which should be of sturdy quality and provides</p>				
--	--	--	--	--	--

		adequate protection of the goods while in transportation : Yes Reports Model approval from Director (Legal Metrology) : Yes Certificate number and date from Director (legal Metrology) Department :IND/--/-- DATED:--/--/20-- Type Test certificate including Environmental conditions from Regional Reference and standard laboratory (RRSL) : Yes RRSL Certificate Number, date and address Calibration from concerned weight and measurement controller ate to be obtained at the time of supplies : Yes Model approval certificate, Type test and calibration certificate to be furnished if asked by buyer at the time of supplies: Yes				
27	Digital BP Apparatus	Type Of SPHYGMOMANOMETERS or Blood Pressure recording unit: Digital Conformity to Indian Standard for SPHYGMOMANOMETERS : NA for digital Measuring device : Automated Type : Integrated ISI MARKED : NA for digital Measurement Method : Oscillometric Display for Digital BP instrument: LCD(Liquid Crystal Digital) Display Range: Pressure measurement (mmHg) : 0 to 300 Range: Pulse measurement for digital with accuracy &plusmn; 5% (per minute) 40 to 200 Accuracy: Pressure measurement (mmHg) : +/- 3 Pressure Detection : Capacitive Pressure sensor Inflation Automatic Pressure Application by pump Deflation and rapid Air Release : Automatic Pressure Release Valve Cuff Range : Medium (22-32)cm Cuff: disinfect able one piece and bladder cuffs, tested according to ISO 105 EO1/ equivalent standard. Operating Temperature range : 0 to 55 degree Celsius Stroage Temperature range : (-10) to 55 degree Celsius Automatic power off if system is idle for 3 minutes : Yes Weight (in gram) : 230 Power Power Source: AA size chargeable battery set Type of AA Size chargeable Battery (AA Size) : Alkaline/ Li-ion WARRANTY, CERTIFICATIONS, TEST REPORTS Availability of Type Test Report covering complete test parameters as per IS3390 latest for Mercurial and IS7652 latest for Aneroid type, from any ILAC/NABL accredited/Central GovLABFor digital type including Environmental test from any ILAC/NABL accredited/Central Gov LAB : Yes CE Marking Certificate for B.P. Apparatus Model. Yes				
		For Digital Sphygmomanometer ,Availability of Type test reports				

		<p>consisting of verification of all the features & functional parameters & environmental tests sequences as under:</p> <p>a) Dry heat test (For 16 hours at a temp of 55degree C in accordance with IS: 9000/part-3/section-3/1977)latest : Yes</p> <p>For Digital Sphygmomanometer Cold test(For 4 hrs at a temp of (-10) degree C in accordance with IS: 9000/part-2/section-3/1977latest e) Damp Heat Cyclic Test (For 2 cycles of 24 hrs at a temp 55 degree C & 95% RH in accordance with IS: 9000/part-5/section-1/1981)latest Note:- The BP Instrument shall be checked for all the parameters before conditioning After completion of the above environmental tests sequence, with a recovery period of 1 to 2 hrs, BP Instrument shall be functional " Yes</p>				
28	Aneroid BP apparatus	<p>CONFIGURATION DETAILS</p> <p>Type Of Sphygmomanometer: Aneroid</p> <p>Design: Corrosion resistant shock proof body, chrome plated metal/ stainless steel pressure control valve</p> <p>Conformity to Indian Standard for SPHYGMOMANOMETERS :IS 7652 latest for Aneroid</p> <p>Measuring device :Mechanical</p> <p>Type : Portable</p> <p>Measurement Method:auscultatory</p> <p>Display for BP instrument:Gauge's background in white color. Graduated scale for ever/ 2mmhg, every 10 units and every 20 units.</p> <p>Range: Pressure measurement (mmHg) : 0 to 300</p> <p>Accuracy:Pressure measurement (mmHg) +/- 3</p> <p>Pressure Detection: NA</p> <p>Deflation and rapid Air Release: Manualsetting of deflation possible up-to 2/3mm Hg/sec. From 260mmHg To 15mm Hg in a maximum deflation time of 10 seconds</p> <p>Cuff Range: (22-32)cm</p> <p>Cuff Type: Nylon straps cuff with pouch, latex bulb with completely chromium plated valve with regulation of vent-hole air by screw valve.</p> <p>Sterilization / Disinfection: The cuff should be disinfect able one piece and bladder cuffs, tested according to ISO 105 EO1/ equivalent standard.</p> <p>Tube : Single rubber tubes used should have an internal diameter of 3 ± 0.5mm and the external diameter should not be less than 8mm</p> <p>Dimensions of the Equipment (mmXmm): NA</p> <p>Dimensions of the Dial screen (mmXmm) : The dial manometer with diameter of 50 mm-60mm</p> <p>Operating Temperature range 0 to 55 degree Celsius</p> <p>Stroage Temperature range (-10) to 55 degree Celsius</p> <p>Weight (in gram) 150</p> <p>WARRANTY, CERTIFICATIONS, TEST REPORTS</p> <p>Availability of Type Test Report covering complete test parameters as per IS7652 latest for Aneroid type, from any ILAC/NABL accredited/Central GovLABFor digital type including Environmental test from any ILAC/NABL accredited/Central Gov LAB: Yes</p> <p>CE/IS Marking Certificate for B.P. Apparatus Model. Yes</p>				
29	Fully Automated Bio-chemistry	Not recommended for HWC PHC as sample load is very minimum				

	Analyzer	<p>compared to the throughput & cost of the System.</p> <p>Semi-automated analyzer can suffice the requirement.</p>				
30	Automated 3Part Differential hematology Analyzer	<p>It is recommended to have 3-part Automated CBC analyzer. where the trained manpower (laboratory technicians are available).</p> <p>Rate contract available at OSMCL which is valid till 13/06/2020.</p>				
31	Urine Analysis(DIPSTICK K)	<p>Specification of DIPSTICK:</p> <p>Dipsticks for urine test for protein and sugar (<i>1 container of 100 strips</i>)</p> <p>Product : Urine Test Reagent Strips</p> <p>Purpose : To determine pathological changes in a patient's urine in standard urinalysis</p> <p>The entire reagent strips should be disposable: Yes</p> <p>Each strip should be stable and ready to use upon removal from the bottle Yes</p> <p>➤ Urine Reagent Strips provide tests for the following parameters:</p> <ol style="list-style-type: none"> 1. Glucose 2. Bilirubin 3. Ketone (Acetoacetic acid) 4. Specific Gravity 5. Blood 6. pH 7. Protein 8. Urobilinogen 9. Nitrite 10. Leukocytes 11. Ascorbic Acid in Urine. <p>➤ The Urine Reagent Strips should be packaged along with a drying agent in a plastic bottle with a cap to provide complete air tight.</p> <p>➤ Each strip should be stable and ready to use upon removal from the bottle.</p> <p>➤ The entire reagent strip should be disposable.</p> <p>➤ Results are obtained by direct comparison of the test strip with the color blocks printed on the bottle label.</p> <p>➤ All the reagent strips should be withstand at a room temperature between 15°-30°C (59°-86°F) and out of direct sunlight.</p> <p>➤ The minimum self-life of the urine strips should be 1year unopened and minimum 3months once it is opened.</p> <p>➤ The required controlled shall be provided along with the strip packet.</p> <p>➤ The strip pack sizes should be of 100 sizes.</p> <p>Quality Standards:</p> <p>➤ Availability of Test Report from Central Govt/NABL/ILAC accredited Lab to prove the conformity to the declared specification (Proof of the same to be submitted to buyer on demand) : YES</p> <p>➤ Availability of Valid Drug License from the appropriate Drug Authority defined under Drugs and Cosmetics Act, 1940(Proof of the same to be submitted to buyer on demand) For Manufacturing Urinalysis Strips, For Sale or Distribution of Urinalysis Strips</p> <p>➤ Drug License Number: To provide</p> <p>➤ Drug License date : To provide</p> <p>➤ ISO 13485/ ICMED 13485 Certified Manufacturer (Proof of the same to be submitted to buyer on demand): YES</p> <p>➤ CE(IVD) /BIS approved</p> <p>➤ Manufacturer should be ISO 13485 certified</p>				
32	Peak Flow Meter	<p>The PeakAir Peak Flow Meter is used to measure your "Peak Expiratory Flow Rate," which is simply the highest speed you can blow air from your lungs.</p> <p>Type: Mechanical Peak Flow Meter</p> <p>Instant readings to help track any changes in the air flow.</p> <p>Dual Flow Range: Accurately measures for both adults and children.</p> <p>Compact, lightweight and portable</p>				

		<p>Built-in flexible three zone management system Easy- to- read numbers Measurement Range 40-800 L/min Accuracy and repeatability assurance Washable</p> <p>Standards EN ISO 23747: 2009 compliance CE Approved/BIS</p> <p>Manufacturer should be ISO 13485 certified</p> <p>WARRANTY: The product is available with 1-year warranty.</p> <p>Peak Flow Meter-1, Reusable Mouthpiece-4 nos(2-Children & 2 Adults), Instruction Manual</p>				
33	Lab Autoclave	<p>Type of Sterilizer: VERTICAL CYLINDRICAL Steam Sterilizer Pressure typeSterilizer Chamber size: 300 mm (Diameter) x 350 mm (Length),with</p> <p>Silicon door gasket (expansion type).</p> <p>Electric load of the unit shall be:2-3 KW.</p> <p>Steam supply to Sterilizer : Steam Generation by electrically heated using immersion heaters wired for operation on 1 phase wire230 Hz AC supply [with immersion heaters having mineral filled sheathed heating elements conforming to IS: 4159/2002 duly BIS marked and complying with safety requirements as per IS : 302-2-201(1992)]</p> <p>it should work with 1 phase 220/230 V 50 Hz AC supply</p> <p>Material: Sterilizer parts; Chamber, Jacket, Door, Door Ring and shall be made of SS 316/SS316L</p> <p>Type of control : Manual</p> <p>Provision of Lock on Steriliser Chamber door: Manual Lock</p> <p>Display facility for Sterilisation Cycle parameters : Manual pressure Gauge Steam & vacuum release valve .</p> <p>Warranty Period: 3 year</p> <p>Pressure,Steel & Spore Test Certificate from a Govt/NABL/ILAC approved testing laboratory shall be furnished by Seller to Buyer for the offered</p> <p>Product should be CE/BIS approved</p> <p>Conformityto standard for electrical safety latest amendment :IEC - 60101</p> <p>Manufacturer should be ISO13485 certified</p>				
34	Refrigerator	200liter capacity may be procured as per availability of service facility .				
35	Incubator	<p>Conformity to Standards: CE/BIS</p> <p>Conformity to standard for for electrical safety latest amendment :IEC - 61010</p> <p>Performance Parameters</p> <p>Purpose Bacteriological incubator are used for the incubation of biological products under controlled conditions</p> <p>Type of Incubator :Bacteriological Incubator</p> <p>Capacity in liters 20</p> <p>Material of Inner Chambers: SS 304</p> <p>Material of Outer Chamber: CRCA (powder coated point) Steel</p> <p>Ambient Temperature in °C: 5°C above ambient to 60 °C (Bacteriological Incubator)</p> <p>Temperature Accuracy in °C: ±0.5°C</p>				

		<p>Temperature Uniformity in °C : ±2°C</p> <p>Insulation : Glass wool</p> <p>Controller : Microprocessor based PID control</p> <p>Digital Display of temperature in °C : Yes</p> <p>Type of Display : LED</p> <p>Type of Temperature sensor : PT100</p> <p>Power supply : 220 V , 50 Hz Single Phase</p> <p>Door: Insulated door and fitted with heavy hinges and should have Toughened glass window Air circulation fan</p> <p>Type of Shelves : Removable</p> <p>Material of Shelves : SS wire mesh</p> <p>Number of shelves :2</p> <p>Size of incubator (L x W x H) in mm x mm x mm:NA</p> <p>Weight in kgs: NA</p> <p>Auto power break off: No</p> <p>Over temperature protection: Yes</p> <p>Safety alarm system : Yes,</p> <p>Instruction manual to be provided : Yes</p> <p>Warranty in Years : 1</p> <p>Standards</p> <p>Manufacturer should have ISO13485 certification: Yes</p> <p>Copies of all certifications and reports to be provided to buyer on demand at time of supplies: Yes</p> <p>Product should be CE/BIS approved</p>				
36	Micro Pipette	<p>Clinical purpose: Pipettes are most appropriate tools for drawing & dispensing of samples for users in the field of clinical diagnostics, control analysis and other areas.</p> <p>Technical Specification:</p> <ul style="list-style-type: none"> ➤ Should have ergonomic design with light & smooth plunger action ➤ Should have soft feel handle grip having both left & right hand operation ➤ Pipette handle should have thermoplastic elastomeric to prevent transfer of body heat to pipette volume during continuous usage ➤ Fully autoclavable: Entire pipette can be steam autoclaved at temp. Of 121 0C ➤ Should have larger & clear 3 digit display giving smaller increment for wider selection of volume ➤ Volume range should be of 10-100µl with increment of 1 µl ➤ Accuracy: 0.8 to 5% ➤ Should have locking mechanism to prevent accidental volume change during pipetting ➤ Should have one hand eject facility ➤ Should have in house clinical, repair and calibration facility ➤ The tip cone should have leak free operation, smooth and light loading operation with choice of using variety of tips. ➤ Should be compatible to universal tip types ➤ Should be available with different color codes. ➤ Warranty: 3years. <p>Quality Standards:</p> <ul style="list-style-type: none"> ➤ Should be USFDA/CE (IVD) approved product ➤ Manufacturers should have ISO 13485 certification for 				

		<p>quality standards</p> <p>➤ Should be applied with individual QC & calibration report according to ISO 8655</p>				
37	Centrifuge-8 Tubes	<p>Performance Parameters</p> <p>Purpose: Separating Serum from blood or for urine sedimentation</p> <p>Centrifuge Design: Bench Top</p> <p>Weight including Default Rotor in kg: NA</p> <p>Auto cut off: Yes</p> <p>Timer provided: Yes</p> <p>Maximum RPM 3000-3500</p> <p>Number of steps for RPM variation :In steps /Continuous</p> <p>Incremental Steps of Variable RPM :300-400</p> <p>Type of Motor: Brushless (without carbons)</p> <p>Maximum Noise level when working at 1 meter distance in db 50</p> <p>Ambient operating Temperature range in Degree centigrade Celsius and humidity range 0 to 50 degree centigrade and Relative humidity 15 to 90%</p> <p>Rotor imbalance diagnostics and automatic stop if required: Yes</p> <p>Facility to ensure that In case Machine is running and lid is opened with manual lock by user , machine shall stop: Yes</p> <p>Maximum Timer Set range 60 minutes</p> <p>Timer settings Provided :interval of 10 seconds up to 1 minute and over 1 minute up to 60 minutes in interval of 1 minute:NA</p> <p>Display :LED /LCD</p> <p>Speed Controller : Microprocessor controlled preprogrammed speed controller</p> <p>Speed Control: Accuracy in case of set speed (Plus minus RPM)</p> <p>Rotor configuration: Swing Out Rotor</p> <p>Rotor Feature with reference to Autoclavability: Autoclavable Rotor @ 121 Degree Celsius,</p> <p>Maximum Volume of the tube testable: 15 ml centrifuges tubes</p> <p>Rotor Size: 8X15ml</p> <p>Power supply :Single phase, 230 V 50 HZ AC Source</p> <p>Power consumption in watts: 500Max</p> <p>Length of Power cord in meters 2</p> <p>The Equipment shall be installed and demonstrated before Acceptance installation and demonstration: Yes</p> <p>Accessories: Spanner to tight or loose Nut of Rotor on Motor shaft head puller, spare rotor nuts ,fuse-2,</p> <p>Warranty in Years 3</p> <p>Material Parameters:</p> <p>Assembly Material for manufacture of Medical Centrifuge: Consists of Die cast Aluminum, Aluminum sheet, Stainless Steel, CRC Steel, Plastic and Rubber</p> <p>Lid Material of Medical Centrifuge : mC.R.C Steel Finish with</p>				

		<p>Powder Coating or stove painting / ABS Plastic of injection molding</p> <p>Body Material of Medical Centrifuge: C.R.C Steel Finish with Powder Coating or stove painting / ABS PLASTIC OF injection molding</p> <p>Centrifuge Bowl material of Medical Centrifuge: 304 Quality Stainless Steel Molded</p> <p>Material of Rotor : Aluminium,</p> <p>Standard:</p> <p>ISO13485 certification for Manufacturer: Yes</p> <p>Copies of all certifications and reports to be provided to buyer on demand at time of supplies Yes</p> <p>Product should be CE/USFDA/BIS approved</p>				
38	Needle Destroyer	<p>GENERAL Purpose :To destroy the used injection needles by instant electro melting process and to cut syringe tips manually using the appliance</p> <p>Type : Portable and electrical</p> <p>Principle of operation : Electro-melting type</p> <p>CONSTRUCTION</p> <p>Suitable sliding tray shall be provided for collection of residue and destroyed syringe tips : yes</p> <p>Material of the Housing /enclosure : ABS Plastic</p> <p>Housing/enclosure shall be moulded type and shock proof: yes</p> <p>Provision to burn the needles and to cut the syringe tips shall be provided in the unit : yes</p> <p>Manual cutter provided in the equipment Made of hardened blade of anti-corrosive SS material</p> <p>Transformer winding : Copper</p> <p>FUNCTIONAL</p> <p>Number of needles of 1mm dia and 80mm length that can be destroyed in continuous operation of 5 minutes 50</p> <p>Sizes of injection needles of all kinds which can be destroyed Dia Ranging from 0.4mm to 1.6mm (26 SWG to 14 SWG) with length 12.5mm to 80mm</p> <p>ELECTRICAL</p> <p>Power supply : 207V to 253V,50Hz AC supply</p> <p>Power ON/OFF switch with indicator shall be provided : YES</p> <p>Rated power(Watts): 60 Watt</p> <p>Shall be provided with fuse and power cord of min 3m length and earthing point : YES</p> <p>Manual of usage instructions : To be supplied by the supplier</p> <p>CERTIFICATIONS</p> <p>Manufacturer should be ISO-13485</p> <p>Comply in general to IS:302Pt-1/1979 to ensure safe and reliable operation in normal use: YES</p> <p>Occupational health and safety certification from approved bodies as per OHSAS 18001-2007 : YES</p> <p>CE /BIS Certification from approved bodies : Yes</p>				

39	Shoulder Wheel Over Head pulley Shoulder ladder	<p>Shoulder wheel, Overhead pulley, Shoulder ladder (set)</p> <p>Clinical use: Used for improving mobility & strength of shoulder girdle & rotator cuff muscles in case of frozen shoulder, paresis & paralysis cases.</p> <p>Technical Specification:</p> <p>A.Shoulder Wheel:</p> <ul style="list-style-type: none"> ➤ It should be a wall mounted one. ➤ The wheel for use of adult and pediatric ➤ The motion arc can be adjustable from 10 to 38 inches by adjusting the handle. ➤ The wheel shall be mounted on a two chrome plated height adjustable (8" to 26") rails. ➤ The resistance can be varied by turning the resistance knob. ➤ The manufacturer should be ISO certified <p>B. Over Head Pulley:</p> <ul style="list-style-type: none"> ➤ Heavy duty, medical grade, shoulders pulley exerciser for physical therapy use. ➤ Over the door metal bracket allows easy one hand setup. ➤ Units with a door strap require two hands, which is difficult with an injured shoulder. ➤ Easily adjustable cord length for any height and for use in both seating and standing position. ➤ The overhead pulley should have the provision of wall mounting. <p>C. Shoulder Ladder:</p> <ul style="list-style-type: none"> ➤ Wood Finger/Shoulder Ladder ➤ 32 vertical finger steps ➤ Solid wood with heavy topcoats ➤ Pre-drilled mounting holes ➤ The manufacturer should be ISO certified. 				
40	Walker	<p>Product type: Frame type, foldable walker</p> <p>Purpose : Walking training device for improving postural alignment while walking, overcoming musculoskeletal impairments and for improving mobility anterior support with less energy consuming</p> <p>Usage : Youth & OLD</p> <p>PRODUCT INFORMATION</p> <p>Posture control possible: Yes</p> <p>Light weight: Yes</p> <p>Wiped clean with a damp cloth: Yes</p> <p>Safety mechanism provided to prevent the user from slipping while walking : Yes</p> <p>Foldable for easy storage and transportation: Yes</p> <p>Anterior support Yes</p>				

		<p>Forearm Supports with latex free flanged hand grips: Yes</p> <p>DIMENSIONS, MATERIAL & WEIGHT</p> <p>Material frame: Stainless steel tubular frame</p> <p>Height of walker with standard wheels (cm) :NA</p> <p>Height of walker with activity wheels (cm): NA</p> <p>Distance between hand grips (cm): 34</p> <p>Frame width (Cm) : 58 to 60</p> <p>Frame length (cm) : 52-59</p> <p>Maximum user weight (kg): 100</p> <p>Frame weight (kg) : 2.5 to 3.0</p> <p>PACKING MODE</p> <p>All the main parts and accessories are Individually bubble wrapped and packed in a strong box to avoid transit damage Yes</p> <p>CERTIFICATIONS & REPORTS</p> <p>Submission of Test Report from Central Government /NABL/ILAC accredited Lab or Manufacture's in house Test Report to prove the conformity to declared specification at the time of supply Yes</p> <p>Certification Available: ISO for manufacturer</p> <p>Product should be CE/ BIS approved</p> <p>WARRANTY:1year</p>				
41	Cervical traction (manual)	<p>Cervical Traction Kitis a complete apparatus designed to provide traction to the cervical and upper dorsal vertebrae.</p> <p>Cervical traction Kit includes the following:</p> <ul style="list-style-type: none"> ➤ Traction Pulley Bracket ➤ Cervical Traction Head Halter ➤ Cervical Traction Spreader Bar ➤ Traction Cord(Nylon rope)5meter. ➤ Traction Weight Bag with weights <p>Controlled continuous traction</p> <p>Easy removal and application</p> <p>Reduced frictional losses</p> <p>Durable and sturdy for a long life and multiple use.</p> <p>Can be used in sleeping posture. Pulley is ergonomically designed to fit any bed side. Soft padded head halter provides a comfortable interface for the traction, is easy to wear and use.</p> <p>Manufacturer should be ISO13485 certified</p> <p>Product should be CE/BIS certified.</p>				
42	Exercise Bicycle	<p>i)Tubular steel frame on properly balanced legs with four rubber tips</p> <p>ii). Fitted with one hard rubber type wheel, standard chain and a socket</p> <p>iii). Seat should be adjustable</p> <p>iv). Should be fitted with a ball bearing resistance roller which permits controlled movement in riding</p> <p>v. LCD/LED display with speedometer.</p>				

		<p>vi. Warranty: 3 Years</p> <p>Manufacturer should be ISO13485 certified</p> <p>Product should be CE/BIS certified</p>				
43	Lumber Traction manual	<p>It strengthening broken bones, immobilization or relieving pressure on the skeletal system.</p> <p>Composition:</p> <p>Universal type</p> <p>Includes pelvic belt, spreader bar & 3 meter long cord.</p> <p>It contains traction water and sand weight bag traction pulley bracket the pelvic traction belt traction cord and 8 kg weight. Manufacturer should be ISO 13485 certified</p> <p>Product should be BIS/CE approved</p>				
44	Gait Training Apparatus/ Parallel bar	<p>General Use: It use for horizontal surface gait training, Coordination training of lower limb in various neurological conditions like paralysis, GBS, Parkinson etc.</p> <p>Technical Specification</p> <p>Parallel bar should be a height adjustable; platform based one which should be very simple for easy accommodation of multiple patients of age group.</p> <p>Handrails should be in single piece, circular in design stainless steel ribbed pipe of minimum diameter 1 1/2".</p> <p>The end of the handrails should be fitted with bumpers.</p> <p>Uprights should be made of heavy gauge square 1 1/2" steel tubes.</p> <p>Parts should be powder coated except the handrail. The handrail should be single piece stainless steel pipe</p> <p>Uprights are to be fitted on a 1 1/2" thick hard plywood base water and termite resistance. Both end of this platform should be curved slope for obstruction free. The platform should be stain finish. Both the end of the platform should have two anti-slip threads for none skidding.</p> <p>Uprights at both ends of platform should be provided for safety of the patient when take turn.</p> <p>Wheels make it safer for patient to turn</p> <p>Minimum dimension of the structure:</p> <p>Length of the bar: 10feet Height of the bar: 25" – 28" Wheel adjustment of the Bar: 26" – 39" Manufacturer should be ISO approved.</p>				
45	Hydro Collator Unit	<p>General Use: A moist heat modality use as a superficial to medium penetration thermotherapy modality. Clinically use for chronic pain management & joint stiffness cases.</p> <p>Technical Specification:</p> <ul style="list-style-type: none"> ➤ Water tank made up of Stainless steel inner & outer Cabin (20-25 Gauge thickness) ➤ High grade thermally insulated ➤ Static/Movable base ➤ Top lid of stainless still, thermally insulated with Fibre handle. ➤ Product Type:Table Top ➤ Dimensions: NA ➤ Tank capacity: 10- L ➤ Temperature range: 50-90 degrees Celsius ➤ Thermostatic temperature control (50 - 90 degree C) - Auto cut-off ➤ Thermal cut-out temperature ➤ Temperature accuracy: +/- 10% ➤ Heat-up time to 90 degree Celsius: in 3 hrs. ➤ Cool-down time from 90 degrees Celsius: 2 hrs. <p>Power supply</p> <p>Mains power: 220-230 V, 50/60 Hz Power consumption: 1000Watt</p> <p>Safety Standard</p> <p>Electrical safety class: Class 1, type B Safety tests: Conforms to IEC 60601-1</p> <p>Quality Standard:</p> <p>The model should be CE/BIS approved. The manufacturer should be ISO13485 certified</p> <p>Accessories must include:</p> <ul style="list-style-type: none"> ➤ Hydro collator Moist Heat Hot Pac : Standard Size-1, Over size-1, Cervical Standard size-1, Hydro collator Knee or Shoulder HotPac-1. 				

		➤ Cotton Towel: 4 pieces.				
46	Traction Bed	<p>Traction Bed – 4 Fold</p> <p>ii. Size : 6 (L) x 2.5 (W) x (H) 3 feet</p> <p>iii. For horizontal cervical and lumbar traction.</p> <p>Should have complete with Storage Shelf & a Flexion Stool</p> <p>Manufacturer should have ISO certified.</p> <p>Product should be CE/BIS Certified</p>				
47	Disposable Gown	<p>Manufactured by An ISO 9001:2008 & CE certified company</p> <p>Surgeon Gown: 40 GSM PPSB</p> <p>Opening both side front and back</p> <p>Gown have attached ties</p> <p>Tissue –Poly –Tissue ,3 play material, size: Gown 30*40"</p>				
48	Disposable Draw Sheet	<p>Disposable made of SMS cloth.</p> <p>Made of good quality interlining fabric with 40 GSM Capacity, Water Proof, Provide with double packing, Size 120cm *120cm,The fabric composite is strongman-microbial ,Anti-Static flame resistant, stain, odor and fluid resistant. The material made of high grade, soft texture, top quality and skin friendly.</p>				

Printing

1	Patient referral cards at PHC Level	<p>Unit : Booklet No. of sheets :50(50 Sheets marked as original +50 Sheets marked as duplicate) with one carbon sheet in each booklet</p> <p>Printing Type (inner sheets): Single side ,Black 1st Sheet of the booklet shall be printed as “Original” and the next sheet shall be printed as “Duplicate ” (for taking carbon copy) and this will be repeated in the remaining sheets of the booklet.,Paper Size :1/4 Demy ,Paper: Original Sheet (White Color)- 80 GSM Maplitho,Brightness:80(Minimum) ,Duplicate Sheet (Yellow Color)54 GSM Maplitho ,Binding Type : Top Pad binding with stapling with ½ pound hard board on back side.Cover Page : 1 cover page on the front (80 GSM Maplitho Single sided Black printing as patient referral Card-PHC), Perforation: Perforation at the top of the pad binding (in original sheet of the booklet only)</p>
2	Patient referral cards at SC Level	<p>Unit : Booklet No. of sheets :50(50 Sheets marked as original +50 Sheets marked as duplicate) with one carbon sheet in each booklet</p> <p>Printing Type (inner sheets): Single side ,Black 1st Sheet of the booklet shall be printed as “Original” and the next sheet shall be printed as “Duplicate ” (for taking carbon copy) and this will be repeated in the remaining sheets of the booklet.,Paper Size :1/4 Demy ,Paper: Original Sheet (White Color)- 80 GSM Maplitho,Brightness:80(Minimum) ,Duplicate Sheet (Yellow Color)54 GSM Maplitho ,Binding Type : Top Pad binding with stapling with ½ pound hard board on back side.Cover Page : 1 cover page on the front (80 GSM Maplitho Single sided Black printing as patient referral Card-Sub Centre), Perforation: Perforation at the top of the pad binding (in original sheet of the booklet only)</p>
3	Filp Book	<p>Unit : Flip Book ,Size:12**8", Total no of fili p sheets : 15 Sheets ,Paper (Flip Sheets): 300 GSM Art Paper (Glosss Finish), Brightness:80 (Minimum)</p> <p>Printing (Flip Sheets): Multi Color both side offset printing ,Lamination : both Top binding with 20 mm wirow , Flip stand cover: Flip stand Cover (1 ½ pound mill board :2 nos connected with each other with base having folding arrangement for hand held as well as table top position) having white Art paper (130 GSM, Brightness ©80 (Min)) Pasiting with wirow binding</p>
4	NCD Durgs Paper Pouch	<p>Unit: Envelop ,Size 4.5**9",Paper: 90 GSM maplitho Paper,Brithtness :77(Minimum),Printing : Single Color Printing (IEC message with logo)</p>
5	Case Record & information Sheet for NTCP	<p>Unit : Booklet , Paper Size :1/4 Demy, Total no of pages :12, Paper 8 Pages :80 GSM Maplitho Paper ,Brightness:77 (Minimum),4 Pages :90 GSM Maplitho Paper ,Brightness :77 (Minimum),</p> <p>Printing:8 pages: Both side single colour printing ,4 pages : Single side multicolour printing ,Bindding : Centre Stitch,Perforation: Perforation at middle of 9th page and side with border at 11th page (in 90 GSM maplitho paper)</p>

Signature of the Bidder

B. FINANCIAL BID - ENVELOPE 'B'

The Bidders have to submit the financial bids in the format prescribed below. This format should be properly filled in, put into a separate envelope, sealed properly and clearly marked as **"Envelope B: Financial Bid for Equipment / Instruments & Printing, Puri"**.

Sl. No.	Name of the Equipment	Category 2. (Equipment / Instruments & Printing, Puri)	Specification	Quantity	Rate/ Price per Quantity	GST	Total Cost (Rs.)
1 (Sl No should be same as sl no of technical specification)	2	3	4	5	6	7	8
Total (in Figure)							
Total (Rs.....) only							

- (i) **Exact quantity will** be intimated before signing of contract.
- (ii) The Total Cost as at Column 8 above is for evaluation purpose only. However, contract will be on per piece per quantity basis.
- (iii) Rates should be inclusive of excise duty / customs duty, packing, forwarding, insurance, transportation charges etc. and exclusive of GST.
- (iv) GST to be quoted (item wise) in separate column No.07.

Note:

Both Envelope A and Envelope B should be put into a bigger envelope and sealed. The envelope should be superscripted on top with the tender reference number and addressed to Chief District Medical & Public Health Officer, Puri . **There is no system of receipt of bids through drop box or by hand. The bids shall be received only through registered post/ speed post/ courier service.** The Tender Inviting Authority shall not be responsible for any postal delay.

Both Cover-A and Cover-B should have an index and page number of all the documents submitted inside that cover.

5. EVALUATION AND SELECTION

- a) The bids shall be evaluated in two stages i.e. technical evaluation and financial evaluation.
- b) The Financial Bids shall be opened of those bidders who will qualify in the technical evaluation.
- c) Financial Bids shall be opened in the presence of the technically qualified bidders/ their representatives, who choose to attend in person at the address given below –

Office of the Chief District Medical & Public Health Officer

PURI, PIN - 752001

Date of Opening of Financial Bids shall be communicated to the technically qualified bidders.

6. AWARD OF CONTRACT:

- a) Contract shall be awarded to the bidder whose bid will be determined to be substantially responsive and who has quoted the lowest evaluated total cost as per individual items.
- b) The Total Cost as at Column 8 above is for evaluation purpose only. However, contract will be per piece per quantity basis.**
- c) In case more than one bidder has quoted the same lowest total cost, the bidder having highest average annual turnover among them will be considered for award of contract.
- d) Intimation of offer for contract will be given by letter/ e-mail. The concerned bidder will have to acknowledge the same and submit the acceptance with the agreement document & Performance Security within 10 days of receipt of such intimation. In case of non-acceptance of agreement, the Tender Inviting Authority may proceed to the next lowest evaluated Bidder.

7. PERFORMANCE SECURITY:

- a) Performance Security acts as a safeguard against unsatisfactory performance or violation of contract agreement by the agency.
- b) 10% of the Purchase order value excluding GST will be retained as Performance Security deposit ~~from~~ the successful Bidder in form of a Demand Draft/ Bank Guarantee which will be returned after 180 days from the date of supply of the items supplied. In case of breach of contract by the agency, the Performance Security shall be forfeited. If the agency duly performs and completes the contract in all respect, the Performance Security shall be returned without any interest on completion of all such obligations under the contract.

SECTION – III

CONDITIONS OF CONTRACT

1. CONTRACT MANAGEMENT

a. Risk and Ownership

i. Liability of damages and losses

1. The agency bears the risk of loss for goods while in transit. The agency shall arrange for appropriate insurance for such goods in transit, the cost of which shall be deemed to be included in the rates set forth in the financial Bid.
2. In case of loss or damage to the goods in transit, amount equivalent to value of such goods will be deducted from the payment to be made to the agency. If value of damaged goods is more than 10 % of total consignment, then additional penalty of 20 % on value of damaged goods will also be levied.
3. The agency shall not be liable for any loss or damage to the extent such is due to a force majeure event, as defined in "FORCE MAJEURE" section of this agreement or an act or default of the Office of the Chief District Medical & Public Health Officer, PURI .

ii. Force Majeure

If through "Force Majeure" (Government embargos, wars, blockages, revolutions, insurrections, mobilizations, strikes, lockouts, riots, other extra ordinary civil disturbances, and/or an act of God) where one or both of the contracting parties are unable to perform their obligations under the terms of this contract; then it shall be considered cancelled and no penalties shall be attached to the parties.

iii. Indemnity

The agency agrees that it shall protect, defend, indemnify and hold harmless to the Office of the Chief District Medical & Public Health Officer, PURI from and against all liabilities, losses, costs, damages, expenses, claims, attorneys' fees and disbursements of any kind or of any nature whatsoever imposed upon the Office of the Chief District Medical & Public Health Officer, PURI , whether incurred directly or indirectly by the Office of the Chief District Medical & Public Health Officer, PURI by virtue of, or in connection with, or arising out of any:

1. Failure of the agency to maintain appropriate licenses to carry out the purposes of this agreement, resulting in the inability to, among other things, ship products for the Office of the Chief District Medical & Public Health Officer, PURI ;
2. Claims made by any employee or agent of the agency or by any operation of the agency related to agency's provisions of transportation services to the Office of the Chief District Medical & Public Health Officer, PURI under the terms of this agreement, including any claim by agency's personnel that they are employees of the Office of the Chief District Medical & Public Health Officer, PURI for any purpose;

b. Penalty & Termination

i. Penalty Clause:

1. The Tender Inviting Authority may also terminate the contract in case of prolonged/ frequent unsatisfactory service, breach of Terms & Conditions and non-compliance by the agency of the instructions given by the Office of the Chief District Medical & Public Health Officer, PURI . **In case of termination of the contract, Performance Security deposited by the agency shall be forfeited.**

ii. Termination

Without prejudice to any other remedies that the Office of the Chief District Medical & Public Health Officer, PURI and its designated agents may have against the agency and its agents, the Office of the Chief District Medical & Public Health Officer, PURI shall have the right at any time by giving notice in writing to the agency to terminate the agreement forthwith in any of the following events:

1. On breach

In case of non-compliance of prescribed quality norms, breach of contractual terms and repeated concerns regarding service quality.

2. On liquidation or insolvency of the Agency

If at any time the agency becomes insolvent or files an application for insolvency or any creditor of his moves the court for adjudicating him as an insolvent or if he is convicted by any court of law, the Office of the Chief District Medical & Public Health Officer, PURI will have the absolute option of terminating the contract forthwith and he shall have no right for damage or compensation on this account.

3. On prevention of performance of duties

The Office of the Chief District Medical & Public Health Officer, PURI shall have the right to terminate this Agreement immediately on notice to agency, if, in the reasonable judgment of the Office of the Chief District Medical & Public Health Officer, PURI , the agency has failed to provide Equipment & Aids with the required standards.

4. On Prejudicial conduct

If the agency is guilty of any conduct which in the opinion of the Office of the Chief District Medical & Public Health Officer, PURI is prejudicial to the interests of the Office of the Chief District Medical & Public Health Officer, PURI .

c. General Provisions

i. Subletting

The agency shall not sublet, transfer or assign the contract or any part thereof, without the previous written approval from the Office of the Chief District Medical & Public Health Officer, PURI . In case the agency contravenes this condition, the Office of the Chief District Medical & Public Health Officer, PURI shall be entitled to place the contract elsewhere at the risk and cost of the agency and all expenses borne on this account shall be recovered from the agency.

ii. Illegal Gratification

1. Any bribe, commission, gift or advantage taken or promised to be taken by or on behalf of the agency or his partner or agent or servant or any one of his authorized representative from the users of the Office of the Chief District Medical & Public Health Officer, PURI shall in addition to any criminal liability that the agency may incur, entitle the Office of the Chief District Medical & Public Health Officer, PURI rescind this contract and all other contracts with him, and to take any other action as may be deemed fit by the Office of the Chief District Medical & Public Health Officer, PURI .
2. The agency shall not lend to or borrow from or have or enter into any monetary dealing or transaction either directly or indirectly with any employee of the Office of the Chief District Medical & Public Health Officer, PURI and if he shall do so, the Office of the Chief District Medical & Public Health Officer, PURI shall be entitled forthwith to rescind the contract.
3. Any question or dispute as to the commission of any offence or compensation payable to the Office of the Chief District Medical & Public Health Officer, PURI under these clauses shall be settled by the designated officer of the Office of the Chief District Medical & Public Health Officer, PURI in such a manner as he shall consider fit and sufficient and his decision shall be final and conclusive.

3. PAYMENT TERMS

1. The agency shall issue an invoice to the Office of the CDM & PHO, PURI for the services rendered under this agreement along with the requisite documents.
2. No advance payment shall be made to the agency. Full payments will be made upon presentation of original signed and stamped Invoices.
3. Payment shall be made through RTGS/ Cheque /e-payment system.

4. RESOLUTION OF DISPUTES

- If dispute or difference of any kind shall arise between the Tender Inviting Authority and the successful bidder in connection with or relating to the contract, the parties shall make every effort to resolve the same amicably by mutual consultations.
- If the parties fail to resolve their disputes or differences by such mutual consultation within twenty-one days of its occurrence, then, unless otherwise provided in the bid document, either Tender Inviting Authority or the successful bidder may give notice to the other party of its intention to commence arbitration, as provided and the applicable arbitration procedure will be as per the Arbitration and Conciliation Act, 1996 of India.

Applicable Law and Jurisdiction of Courts

- The contract shall be governed by and interpreted in accordance with the laws of India.
- All disputes arising out of this bid will be subject to the jurisdiction of courts of law in PURI / High Court of Odisha.

ENCLOSED IN SEPARATE SHEET

ANNEXURE- II

TURNOVER CERTIFICATE

I hereby certify that M/s. _____ (Name & address _____) is having the following annual turnover and the statement is true and correct -

Sl. No.	Financial Year	Turnover
1.	2016-17	Rs. _____ (Rupees _____) only
2.	2017-18	Rs. _____ (Rupees _____) only
3.	2018-19	Rs. _____ (Rupees _____) only

Signature of the Bidder:

Date:

Signature of Auditor/ Chartered Accountant

Membership No.:

Seal:

Declaration

To

The Chief District Medical & Public Health Officer

PURI

I/We_____have carefully read the tender document and confirm my/ our eligibility as required. I/ We solemnly declare that I/ we have never been convicted by any court of law or blacklisted/ debarred on financial/administrative/techno-legal ground by any appropriate authority. By submitting this bid, I/ we assure my/ our agreement to the terms and conditions of tender and will abide by the same if contract is awarded to me/ us.

Currently running criminal/civil suits against my/ our firm are:

(Write "NIL" if no court case pending otherwise give the list and enclose case details)

Signature of the bidder:

With seal

NOTE: To be given in a Rs. 10/- stamp paper.

General Details of the Bidder

Sl. No	Particulars	Details
1.	Name of the Organization	
2.	Nature of Business	
3.	Year of Inception	
4.	Name and Details of the Proprietor, Firm/ Director of Company/ Organization/ Agency/Supplier/Distributor Mobile No: Email Address:	
5.	Full Address of Regd. Office Telephone Number Fax. No. E-Mail Address	
6.	Website of the Organization	
7.	PAN	

AGREEMENT

This agreement is made on this _____ day of _____, 2019.

BETWEEN

Office of the Chief District Medical & Public Health Officer, PURI, having its registered office at PURI (Herein after "The 1st Party")

AND

M/s. -----, having its registered office at -----

--, represented through its -----, duly authorized for the purpose (Herein after "The 2nd Party").

Whereas the 1st Party has invited tender for "**Equipment / Instruments & Printing for Model PHC – HWCs , PURI**" vide tender reference no. _____ dt. _____ .2020. The 2nd Party submitted its bid as per tender requirement and being the L1 bidder, the 1st Party has finalized the tender in favour of the 2nd Party for supplying **Equipment / Instruments & Printing for Model PHC – HWCs , PURI** at the prices/rates indicated against various components (herein after "Contract Price") on the terms and condition set forth in this agreement. The letter of award was issued vide LOA No. _____ dated _____ inviting to execute the contract. And whereas the 2nd Party has deposited the Performance Security of Rs. _____ vide a Demand Draft/ Bank Guarantee No. _____ dt. _____.

NOW THIS AGREEMENT WITNESSES AS FOLLOWS:

1. The following documents shall be deemed to form and be read and construed as integral part of this Agreement, viz.:
 - a) Conditions of Contract
 - b) Submissions and Declaration as part of the Proposal submitted
 - c) Notification of Award issued by the Office of the CDM&PHO, PURI
2. In consideration of the payments to be made by the 1st Party to the 2nd Party, the 2nd Party hereby covenants with the 1st Party to provide the agreed services uninterrupted in all respects as per the provisions of this Contract.
3. The 1st Party hereby covenants to pay the 2nd Party in consideration of the provision of the agreed services, the Contract Price or such other sum as may become payable under the provisions of the Contract at the times and in the manner prescribed in the Contract.

4. TENURE OF CONTRACT

- a. The contract is for a period of one (1) years from the date of signing of contract.

- b. In the event of the contract being terminated by the 1st Party before the completion of the contract period for reasons of unsatisfactory performance or breach of contractual conditions, the Performance Security of the 2nd Party shall be liable to be forfeited.

For and on behalf of the Agency

Chief District Medical & Public Health Officer, PURI

Authorized Signatory

Authorized Signatory

(Name with Signature)

(Name with Signature)

Date:

Date:

1. Witness

1. Witness

2. Witness

2. Witness

ANNEXURE VI

MODEL BANK GUARANTEE FORMAT FOR PERFORMANCE SECURITY

[Refer Para 7]

To

The Chief District Medical & Public Health Officer

PURI

WHEREAS (Name
and

Address of the Service Provider) (here in after called "the Service Provider") has undertaken, in pursuance of Tender Reference No. dated..... for engagement of commercial transport vehicles for **Supplying Equipment / Instruments & Printing for Model PHC –HWCs , PURI** (herein after called "the contract").

AND WHEREAS it has been stipulated by you in the said contract that the Service Provider shall furnish you with a bank guarantee by a scheduled commercial bank recognized by you for the sum specified therein as security for compliance with its obligation in accordance with the contract.

AND WHEREAS we have agreed to give the Service Provider such a bank guarantee;

NOW THEREFORE we hereby affirm that we are guarantors and responsible to you, on behalf of the Service Provider, up to a total of(amount of the guarantee in words and figures), and we undertake to pay you, upon your first written demand declaring the Service Provider to be in default under the contract and without cavil or argument, any sum or sums within the limits of (amount of guarantee) as aforesaid, without your needing to prove or to show grounds or reasons for your demand or the sum specified therein.

We hereby waive the necessity of your demanding the said debt from the Service Provider before presenting us with the demand.

We further agree that no change or addition to or other modification of the terms of the contract to be Performed there under or of any of the contract documents which may be made between you and the Service Provider shall in any way release us from any liability under this guarantee and we hereby waive notice of any such change, addition or modification.

This guarantee shall be valid until theday of.....

We theBranch..... undertake not to revoke the guarantee during its currency except with the previous consent of the Office of the Collector & District Magistrate, PURI in writing.

WeBranch..... further agree that a mere demand by the Office of the Chief District Medical & Public Health Officer, PURI is sufficient for us to pay the amount covered by the Bank Guarantee without reference to the said Agency and protest by the said Agency cannot be a valid ground for us to decline payment to the Office of the Chief District Medical & Public Health Officer, PURI .

(Signature of the authorized officer of the Bank)

Name and designation of the Officer

Seal, name & address of the Branch of the Bank

