

Drugs & M.C for Cataract surgery

Sl No	Name of the items	Rate	GST	Total Cost
1	Dark Glass			
2	IOL(18,19,20,21,22,23)			
3	SICS Blade(Extension 502 MM Blunt tip)			
4	SICS Blade(Keratome 2.8 MM Point tip)			
5	SICS Blade(Crescent Knife)			
6	Inj Hynidase			
7	Swab Stick			
8	Blue Raxis			
9	Tropicamide+Phynilephrine Eye Drop			
10	Xylocaine 4%			
11	Paracaine Eye Drop			
12	Inj Pilocarpine			
13	Inj Visco			
14	Moxifloxacin+Dexamethasone Eye Drop			
15	Hospal OT Solution			
16	Cyclopentolate Eye Drop			
17	Acetone			
18	Artery Forcep(8")			
19	Limbs Foprcep(tooth)			
20	Simcoi Cannula(20 G)			
21	Eye Speculum(wire speculum)			
22	Corneal Scissor(curved)			
23	Conjunctival Scissor			
24	Vanous Scissor			
25	Balansed Salt Solution			
26	Lens Holding Forcep			
27	Suture Tying Forcep			
28	Superior Rector Holding Forcep			
29	Swiskey's Hook			
30	IOL Dialer			
31	8-0 Nylon			
32	5-0 Silk			
33	5-0 vicryl			
34	Binacullar Loope			

Specification for Instruments

The Instrument Steel Should comply to EN ISO 7153 standard for stainless steel used for surgical instruments and ISO 13402 standard,determination resistance against autoclaving,corrosion and thermal exposure,necessary certificate should be furnished.

The Instrument must be vaccum hardened, in order to receive the necessary hardness tenacity and corrosion resistance.

Stainless steel material grade should be AISI 410 and 420 certified and should support by test report.

All instrument should be provided with 2 yrs replacement warranty.

Chief District Medical & Public Health Officer, Puri