

**CORRIGENDUM**

No: 1619

Date:30.05.2026

In partial modification to the Tender Advertisement No: 1551 on date 20.05.2026 which was published through <https://puri.odisha.gov.in/> and two odia news paper (SAMAJ & SAMBAD) on dtd. 21.05.2026. Interested bidders are requested to read that Annexure-A, The Technical Specification of calibrated blood Collection Drape. Sl. No-2, Sl. No-9 are modified. **The revised details of. Sl. No-2, Sl. No-9 and three additional clauses added are available in the district website <https://puri.odisha.gov.in/>.** All other conditions are remain same. If any updates from state, it will be discussed and finalized, submission of tender document extended to on or before 11.06.2026 (5.00PM) to the undersigned through speed post/by registered post/courier superscribing on the top left hand corner of the envelop.

*Asst. PHO 30.5.2026*  
**CDM&PHO,Puri**

## CORRIGENDUM

No. 1619

Date. 30.05.26

The following amendments/clarifications may please be noted and treated as part of the tender document:

Sl. No.	Feature/ Terms & Conditions	Specification/ Existing Clause / Section	Corrigendum / Amended Clause
1	Dimensions	Rectangular Under Buttock Sheet: Width: 80 to 90 cm and Length: 55 to 75 cm. The rectangular sheet should have straps of 2 to 3 inches width, which can be used to tie around the abdomen for securing the sheet properly.	Rectangular Under Buttock Sheet: Width: 80 to 90 cm and Length: 65 to 75 cm. The rectangular sheet should have straps of 2 to 3 inches width, which can be used to tie around the abdomen for securing the sheet properly.
2	Material Composition	<b>Essential:</b> Main Rectangular Sheet: Non-woven polypropylene laminated with polyethylene, minimum 60 GSM, without any absorbent layering. Pouch: Transparent medical-grade PVC or polyethylene.	<b>Essential:</b> Main Rectangular Sheet: Non-woven polypropylene laminated with polyethylene, minimum 60 GSM, Not to have any absorbent layering. Pouch: Transparent PVC or polyethylene (Medical Grade), minimum 75 micron thickness. <b>Desirable:</b> Bridgeable material for the main sheet and pouch.
3	Eligibility Criteria & Documents to be Submitted in Bid	-	Sample of the quoted product/equipment must be submitted before the bid end date for technical evaluation and verification purposes.
4	Eligibility Criteria & Documents to be Submitted in Bid	-	All the specifications mentioned in the tender document must be clearly reflected and supported in the test report. The test report must be issued by a Government Accredited Laboratory / NABL Accredited Laboratory and should comply with the tender specifications.
5	Eligibility Criteria & Documents to be Submitted in Bid	-	The bidder must submit detailed catalogue and other supporting documents of the product indicating detailed specifications and features of the quoted product.

All other terms and conditions of the tender shall remain unchanged.

*Asatpa 30.5.26*  
CDM&PHO, Puri